

Certificate of Analysis

Product Information

Product Name:	PlexPCR® Titanium Core Reagents	
Catalogue Number:	6001021	
Product Description:	The Plex PCR® Titanium Core Reagents are real-time polymerase chain reaction (qPCR) reagents for detection of DNA.	
Lot Number:	22040075	
Expiry Date:	2025/01/16	
Storage Conditions:	-15°C to -25°C	

Test method:

The PlexPCR® Titanium Core Reagents were tested by extracting internal control cell DNA and amplifying it in a qPCR reaction on the ABI 7500 FAST Dx thermocycler.

Results:

Test Samples	Specification	Result
Internal Control Cell DNA	Cq values must fall within a set range.	PASS

The Finished products were tested according to standard Quality Control methods. The results were verified against the design specifications, acceptance criteria and reviewed and approved by the Quality Department.

Quality Assurance Signature and Date:

QA Approval	<u>Ivy Chau</u> Ivy Chau (Apr 8, 2023 20:49 GMT+10)
-------------	--